

Underwriters Laboratories (UL LLC) Safety Certification Report



Model: LWFA2E120-092SS-07
LWFA2E120-092SU-08

Device Description: Electrical Fan

Applicant: Ningbo Longwell Electric Technology Co., Ltd
Xuxiang River 100
Yuyao, Zhejiang Sheng, 315400 China

Manufacturer: Same as Applicant

Manufacturing Facility(ies): Same as Applicant

Report No.: E552586-D1000-1/A0/C0-UL

Report (Re)Issue Date: 2026-02-12

Base Standard(s): AAMI ES60601-1:2005,ES60601-1:2005/AMD1 1:2012 , ES60601-1:2005/AMD2:2021, CAN/CSA-C22.2 No. 60601-1:08, CAN/CSA-C22.2 No. 60601-1:14 (including amendment 1) and Amendment 2:2022 (MOD) to CAN/CSA-C22.2 No. 60601-1:14

Additional Standards: N/A

Report Types: This report consists of the following report types:
[Yes] US Certification (UL Unlisted Component)
[Yes] CAN Certification (cUL Unlisted Component)

This report covers the Safety evaluation of the referenced model(s) according to the standard(s) specified above.

Description

UL TEST REPORT AND PROCEDURE

Standard:	AAMI ES60601-1:2005,ES60601-1:2005/AMD1 1:2012 , ES60601-1:2005/AMD2:2021, CAN/CSA-C22.2 No. 60601-1:08, CAN/CSA-C22.2 No. 60601-1:14 (including amendment 1) and Amendment 2:2022 (MOD) to CAN/CSA-C22.2 No. 60601-1:14
Certification Type:	Unlisted Component
CCN:	MDAF33 / MDAF39
Complementary CCNs:	MDAF9
Product:	Electrical Fan
Model:	LWFA2E120-092SS-07 LWFA2E120-092SU-08
Rating:	LWFA2E120-092SS-07: 230Vac, 50Hz LWFA2E120-092SU-08: 115Vac, 60Hz
Applicant Name and Address:	Ningbo Longwell Electric Technology Co., Ltd Xuxiang River 100 Yuyao, Zhejiang Sheng, 315400, China

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

UL authorizes the applicant to reproduce the latest pages of the referenced Test Report consisting of the first page of the Specific Technical Criteria through to the end of the Conditions of Acceptability as applicable.

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

Prepared by: Lars Groennegaard,
Handler

Reviewed by: Mona H. Nielsen, Reviewer